

Correction for Interferogram Asymmetry in Hyperspectral Signals

Mark Fitzgerald (University of Colorado, Denver)
perfabe@email.msn.com

Abstract

The Michelson interferometer can be used to collect an infrared signature over a wide band of wavelengths. Light reflection within the sensor material leads to destructive interference in the signal that is produced. However, the structure of the interference can be estimated, and phase corrections based on these estimates can be implemented to reconstruct the symmetric interferogram.